


HE/LX[®] (Holter) Analysis - Software Features Matrix

Software Feature / Arrhythmia	Software Level			
	Pro	Enhanced+	Enhanced	Basic
International Languages	✓	✓	✓	✓
1 Strip Per Hour	✓	✓	✓	✓
Bi-Directional Scanning	✓	✓	✓	✓
Diary Entry	✓	✓	✓	✓
Event Marker	✓	✓	✓	✓
Label Strips	✓	✓	✓	✓limited
Maximum and Minimum HR	✓	✓	✓	✓
Disable SVPBs	✓	✓	✓	✓
Page Mode	✓	✓	✓	✓
Six Minute Walk Assessment	✓	✓	✓	✓
SpO ₂	✓ ²	✓ ²	✓ ²	✓ ²
Calibration Strips	✓	✓	✓	
Configurations	✓	✓	✓	
Critical Events	✓	✓	✓	
Failure To Capture	✓	✓	✓	
Failure To Sense	✓	✓	✓	
Inhibition	✓	✓	✓	
Insert Beat	✓	✓	✓	
Interval Editor	✓	✓	✓	
Pause	✓	✓	✓	
QTc – 4 choices for calculating	✓ ¹	✓ ¹	✓ ¹	
Saved Strips	✓	✓	✓	
ST Baseline	✓	✓	✓	
ST Depressions and Elevations	✓	✓	✓	
ST Deviation (Maximum)	✓	✓	✓	
SVPB, SVPB Pair, SVT	✓	✓	✓	
Time Domain HRV	✓	✓	✓	
True 12-Lead	✓ ¹	✓ ¹	✓ ¹	
VPB, VPB Pair, VTAC	✓	✓	✓	
AF Trend	✓	✓		
AF Parameters	✓	✓		
Receive ECG data/Transmit report files	✓	✓		
Review Bins	✓	✓		
Templates	✓	✓		
Beats	✓	✓		
AF Events and AF Strips	✓			
Bigeminy	✓			
Bradycardia and Tachycardia	✓			
Fastest and Longest AF	✓			
Fastest and Longest SVT	✓			
Fastest and Longest VTAC	✓			
Hospital Information System File	✓			
Irregular and Longest R-R	✓			
Quadrigeminy	✓			
Spectral Analysis	✓			
Spectral Plots & Tables	✓			
ST End	✓			
ST Maximum HR	✓			
ST Onset	✓			
Trigeminy	✓			
User Login and Audit available	✓			

1 – DR181 recorder w/ 12-Lead required; 2 – DR181 w/ purchase of SpO2 cable required

Industry Leading 3-Year Warranty



NorthEast Monitoring, Inc.
advancing Holter technology™

Northeast Monitoring, Inc.

Distribuido por
Systemetics Spain S.L.


Tel. +34609888805

systemetics@systemetics.cat

HE/LX[®] (Holter) Analysis - Report Features Matrix

Report Feature	Software Level			
	PRO	ENHANCED+	ENHANCED	BASIC
Comments page	✓	✓	✓	✓
Full disclosure - 30 or 60 minutes/page	✓	✓	✓	✓
Full disclosure strips	✓	✓	✓	✓
Full-sized strips	✓	✓	✓	✓
International languages	✓	✓	✓	✓
List of diary events	✓	✓	✓	✓
Oxy and HR trend - 24 hours	✓ ²	✓ ²	✓ ²	✓ ²
Oxy and HR trend - 2 hours	✓ ²	✓ ²	✓ ²	✓ ²
Oxy and heart rate summary	✓ ²	✓ ²	✓ ²	✓ ²
Oxy values with full disclosure	✓ ²	✓ ²	✓ ²	✓ ²
Oxy trend with full disclosure	✓ ²	✓ ²	✓ ²	✓ ²
Patient information summary: 6 options	✓	✓	✓	3 types
Review and edit front page	✓	✓	✓	✓
Six-Minute-Walk Assessment	✓	✓	✓	✓
12-Lead Strips				
12-lead strips - 25 or 50 mm/sec	✓ ¹	✓ ¹	✓ ¹	
General profile and trends	✓	✓	✓	
HRV time summary	✓	✓	✓	
Paced data information / heart rate trend	✓	✓	✓	
Saved strips list	✓	✓	✓	
ST episodes	✓	✓	✓	
Time domain table	✓	✓	✓	
Strips and Bins				
Half-sized Strips	✓	✓		
Normal bins	✓	✓		
Paced bins	✓	✓		
Ventricular bins	✓	✓		
Advanced Analysis				
Color reports optional	✓			
12-lead table	✓ ¹			
12-lead trends	✓ ¹			
3-D plot "waterfall graph" (spectral)	✓			
Bigeminy Table	✓			
Bradycardia/Tachycardia Table	✓			
Circadian plots (spectral)	✓			
Critical events	✓			
Frequency domain table (spectral)	✓			
Hourly rhythm page	✓			
Normalized frequent domain (spectral)	✓			
Paced interval histograms	✓			
Paced summary information	✓			
Expanded HR and ST trends	✓			
Settings / Scanning criteria	✓			
Spectral HRV (frequency domain)	✓			
ST analysis triaxial graph	✓ ¹			
Supraventricular summary	✓			
Ventricular summary	✓			
Audit file available	✓			

1 – DR181 recorder w/ 12-Lead required; 2 – DR181 w/ purchase of SpO2 cable required



NorthEast Monitoring, Inc.
advancing Holter technology™



NB: In the U.S.A., Federal Law restricts devices to sale by or on the order of a physician. Not for in vitro diagnostic use. Not intended to replace real-time telemetry monitoring for patients suspected of having life-threatening arrhythmias.
systemetics@systemetics.cat